

YOUR REF. NO.:
OF DATE:
OUR REF. NO.: 009/2016-6015

HANDLED BY: Jindřich Bílek
PHONE: +420 545 555 717
E-MAIL: jbilek2@cmi.cz

DATE: 2016-07-22

Euromag International SRL
Raffaele Zocchi
Via Torino 3-35035 - Mestrino
Padova - Italy

Issuing of addition 1 to OIML Certificate of conformity

Dear Sir,

Czech Metrology Institute as OIML issuing authority confirms that water meters of company Euromag International SRL type MC608A/MUT2200/EL with PTFE lining passed the testing according OIML R49:2013 with positive result for measuring in vertical and reverse position. Tests were finished in June 2016.

The issuing of the addition 1 to OIML certificate of conformity is just administration procedure and will be completed as soon as possible.

Best Regards

Ing. František Staněk Ph.D.
Director for legal metrology



Český metrologický institut
Oblastní inspektorát Brno
Okružní 31
638 00 Brno
-18-